

HF Calibrator

For MP100 HF Monitor
Quick Start Guide



mPower Electronics Inc.

3046 Scott Blvd. Santa Clara, CA 95054
www.mpowerinc.com
info@mpowerinc.com

v1.0 Aug. 2023

Specifications

| | |
|-----------------------|--|
| Concentrations | 1-15 ppm |
| Accuracy | ±15% (new) |
| Operating Life | 200 calibrations or 6 months, whichever is first |
| Temperature | 15-30°C (59-86°F) |
| Storage Temp. | 10-30°C (50-86°F) |
| Pressure | 1±0.1 atm |
| Humidity | 15-90% non-condensing |
| Weight | 220 g |
| Height | 15.7 cm (6.2 in.) |
| Diameter | 11.4 cm (4.5 in.) |

Overview

The HF Generator has a reservoir that contains a sorbent material onto which HF is impregnated. It creates a constant HF gas concentration in the 1-15 ppm range, depending on the temperature. It is designed for bump test and calibration of diffusion gas monitors and cannot be used for pumped instruments because there is no flowing gas. The outlet fitting is specifically made to fit mPower UNI gas monitors.

Power Connection

Power the unit with 3 AAA alkaline batteries or plug a USB-C cable into the USB slot and connect to a 5VDC power supply.

Operation

1. Allow the unit at least at least 30 minutes at ambient temperature for the concentration reading to stabilize.
2. Adjust the UNI Cal Set value to match the concentration reading on the calibrator.
3. Place a piece of foam about 1 cm (3/8") thick on the battery compartment.
4. Unscrew the calibrator cap and set aside.
5. Start the UNI calibration timing and quickly place the UNI face down with the sensor head onto the HF gas opening and the display supported by the 1-cm foam, and open the valve.

Operation (continued)

6. Start a stopwatch to measure when the calibration time is complete before removing the UNI. Add one minute if it is desired to see the final span reading, to allow the UNI display to automatically revert to the gas reading. Read immediately, as this value will drop quickly once the UNI is removed from the gas source.
7. Close the valve and screw the cap back on tightly.
8. NOTE: Handle the unit gently; do not shake as this may cause damage and will not help speed up the concentration equilibration process.

